

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2928</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Hero Construction</u>	DataFile :	<u>PP074600.D</u>

		Sample				Rec		RPD		Limits	
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2915-03MSD	Client Sample ID:		BRIDGE-EASTMSD							
(Column 1)											
AR1016	196.9	0	193	ug/kg	98		3		55	146	15
AR1260	196.9	0	176	ug/kg	89		3		54	119	15
Lab Sample ID:	Q2915-03MSD	Client Sample ID:		BRIDGE-EASTMSD							
(Column 2)											
AR1016	196.9	0	192	ug/kg	98		4		55	146	15
AR1260	196.9	0	166	ug/kg	84		4		54	119	15