



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2928</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Hero Construction</u>	Datafile :	<u>PP074583.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169359BS (Column 1)	AR1016	166.6	156	ug/kg	94				71	120	
	AR1260	166.6	151	ug/kg	91				65	130	
PB169359BS (Column 2)	AR1016	166.6	140	ug/kg	84				71	120	
	AR1260	166.6	139	ug/kg	83				65	130	