

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Lab Code: ACE

Contract: HERO01
SDG NO.: Q2928

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/18/2025 08/18/2025
Client Sample No. (PEM): PEM - PD089962.D Date Analyzed: 08/18/2025
Lab Sample No.(PEM): PEM Time Analyzed: 10:57

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.067	8.970	9.170	19.240	20.000	-3.8
Tetrachloro-m-xylene	3.547	3.500	3.600	18.000	20.000	-10.0
alpha-BHC	3.996	3.950	4.050	8.460	10.000	-15.4
beta-BHC	4.514	4.460	4.560	9.550	10.000	-4.5
gamma-BHC (Lindane)	4.327	4.280	4.380	8.800	10.000	-12.0
Endrin	6.571	6.500	6.640	46.440	50.000	-7.1
4,4'-DDT	7.017	6.950	7.090	92.340	100.000	-7.7
Methoxychlor	7.490	7.420	7.560	216.780	250.000	-13.3

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/18/2025 08/18/2025
Client Sample No. (PEM): PEM - PD089962.D Date Analyzed: 08/18/2025
Lab Sample No.(PEM): PEM Time Analyzed: 10:57

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.065	7.960	8.170	18.090	20.000	-9.6
Tetrachloro-m-xylene	2.878	2.830	2.930	18.480	20.000	-7.6
alpha-BHC	3.390	3.340	3.440	9.940	10.000	-0.6
beta-BHC	4.022	3.970	4.070	10.050	10.000	0.5
gamma-BHC (Lindane)	3.726	3.680	3.780	9.780	10.000	-2.2
Endrin	5.784	5.710	5.850	44.880	50.000	-10.2
4,4'-DDT	6.177	6.110	6.250	86.730	100.000	-13.3
Methoxychlor	6.748	6.680	6.820	185.620	250.000	-25.8