

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082225\
 Data File : PD090008.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Aug 2025 14:47
 Operator : AR\AJ
 Sample : PB169360BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB169360BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 01:52:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 19 13:16:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.877	55743176	478.1E6	18.994	21.681
28)	SA Decachlor...	9.070	8.065	92258350	584.7E6	21.376	22.586
Target Compounds							
2)	A alpha-BHC	3.998	3.389	279.0E6	1615.2E6	46.490	47.593
3)	MA gamma-BHC...	4.329	3.725	266.3E6	1491.8E6	46.501	47.858
4)	MA Heptachlor	4.927	4.077	266.1E6	1478.7E6	47.439	47.406
5)	MB Aldrin	5.269	4.363	253.0E6	1461.8E6	46.722	47.997
6)	B beta-BHC	4.515	4.021	100.7E6	631.4E6	44.895	47.365
7)	B delta-BHC	4.763	4.257	262.3E6	1510.0E6	50.658	47.861
8)	B Heptachlo...	5.689	4.867	228.3E6	1332.0E6	46.529	48.096
9)	A Endosulfan I	6.072	5.241	216.4E6	1264.0E6	46.881	48.203
10)	B gamma-Chl...	5.944	5.119	229.0E6	1442.1E6	46.819	48.702
11)	B alpha-Chl...	6.025	5.184	229.3E6	1371.7E6	46.588	48.064
12)	B 4,4'-DDE	6.194	5.369	206.6E6	1384.2E6	47.088	48.024
13)	MA Dieldrin	6.345	5.506	231.4E6	1400.3E6	47.274	47.814
14)	MA Endrin	6.572	5.783	193.2E6	1293.6E6	46.617	48.574
15)	B Endosulfa...	6.785	6.075	187.2E6	1217.4E6	44.369	48.129
16)	A 4,4'-DDD	6.704	5.923	163.6E6	1162.1E6	47.304	48.334
17)	MA 4,4'-DDT	7.019	6.177	179.2E6	1216.3E6	47.046	48.517
18)	B Endrin al...	6.913	6.252	143.2E6	902.6E6	49.961	52.416
19)	B Endosulfa...	7.148	6.476	182.3E6	1173.6E6	46.527	48.137
20)	A Methoxychlor	7.492	6.747	91275999	595.4E6	45.510	46.171
21)	B Endrin ke...	7.629	6.985	196.0E6	1300.6E6	47.291	48.811
22)	Mirex	8.112	7.178	143.1E6	993.8E6	45.352	48.114

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082225\
Data File : PD090008.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Aug 2025 14:47
Operator : AR\AJ
Sample : PB169360BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB169360BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 01:52:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
Quant Title : GC Extractables
QLast Update : Tue Aug 19 13:16:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

