

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2928</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Hero Construction</u>	DataFile :	<u>PD090021.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits		
			Result				Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2915-01MSD (Column 1)	Client Sample ID:		BRIDGE-WESTMSD								
	alpha-BHC	18.88	0	17.7	ug/kg	94		3		60	144	20
	beta-BHC	18.88	0	17.1	ug/kg	91		4		54	143	20
	delta-BHC	18.88	0	15.5	ug/kg	82		4		29	151	20
	gamma-BHC (Lindane)	18.88	0	16.9	ug/kg	90		3		61	140	20
	Heptachlor	18.88	0	15.8	ug/kg	84		4		63	135	20
	Aldrin	18.88	0	17.6	ug/kg	93		3		49	139	20
	Heptachlor epoxide	18.88	0	17.1	ug/kg	91		4		41	156	20
	Endosulfan I	18.88	0	17.0	ug/kg	90		5		56	142	20
	Dieldrin	18.88	0	16.8	ug/kg	89		5		47	161	20
	4,4'-DDE	18.88	0	17.7	ug/kg	94		8		55	136	20
	Endrin	18.88	0	16.6	ug/kg	88		5		57	139	20
	Endosulfan II	18.88	0	16.0	ug/kg	85		5		40	163	20
	4,4'-DDD	18.88	0	17.1	ug/kg	91		6		47	163	20
	Endosulfan sulfate	18.88	0	14.8	ug/kg	78		1		62	139	20
	4,4'-DDT	18.88	0	13.1	ug/kg	69		6		51	146	20
	Methoxychlor	18.88	0	12.7	ug/kg	67		9		54	136	20
	Endrin ketone	18.88	0	14.9	ug/kg	79		4		60	129	20
	Endrin aldehyde	18.88	0	15.9	ug/kg	84		10		59	132	20
	alpha-Chlordane	18.88	0	16.9	ug/kg	90		6		39	166	20
	gamma-Chlordane	18.88	0	17.2	ug/kg	91		4		44	175	20
Lab Sample ID:	Q2915-01MSD (Column 2)	Client Sample ID:		BRIDGE-WESTMSD								
	alpha-BHC	18.88	0	17.3	ug/kg	92		2		60	144	20
	beta-BHC	18.88	0	17.6	ug/kg	93		2		54	143	20
	delta-BHC	18.88	0	14.5	ug/kg	77		3		29	151	20
	gamma-BHC (Lindane)	18.88	0	16.9	ug/kg	90		2		61	140	20
	Heptachlor	18.88	0	15.8	ug/kg	84		1		63	135	20
	Aldrin	18.88	0	17.7	ug/kg	94		3		49	139	20
	Heptachlor epoxide	18.88	0	17.4	ug/kg	92		2		41	156	20
	Endosulfan I	18.88	0	17.3	ug/kg	92		3		56	142	20
	Dieldrin	18.88	0	17.1	ug/kg	91		3		47	161	20
	4,4'-DDE	18.88	0	17.5	ug/kg	93		3		55	136	20
	Endrin	18.88	0	17.1	ug/kg	91		4		57	139	20
	Endosulfan II	18.88	0	16.4	ug/kg	87		4		40	163	20
	4,4'-DDD	18.88	0	16.3	ug/kg	86		4		47	163	20
	Endosulfan sulfate	18.88	0	14.9	ug/kg	79		3		62	139	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2928</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Hero Construction</u>	DataFile :	<u>PD090021.D</u>

Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
		Result	Result			Qual	RPD	Qual	Low	High	RPD
4,4'-DDT	18.88	0	14.4	ug/kg	76		3		51	146	20
Methoxychlor	18.88	0	13.5	ug/kg	72		3		54	136	20
Endrin ketone	18.88	0	14.6	ug/kg	77		4		60	129	20
Endrin aldehyde	18.88	0	16.5	ug/kg	87		4		59	132	20
alpha-Chlordane	18.88	0	17.4	ug/kg	92		2		39	166	20
gamma-Chlordane	18.88	0	17.6	ug/kg	93		2		44	175	20