

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081425\
 Data File : BP025400.D
 Acq On : 14 Aug 2025 18:05
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP081425

Quant Time: Aug 14 18:24:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.472	0.399	15.5	99	0.00
3	Pyridine	1.291	1.159	10.2	105	0.00
4	n-Nitrosodimethylamine	0.532	0.507	4.7	109	0.00
5 S	2-Fluorophenol	1.204	1.045	13.2	99	0.00
6	Aniline	2.080	1.984	4.6	108	0.00
7 S	Phenol-d6	1.544	1.383	10.4	101	0.00
8	2-Chlorophenol	1.359	1.308	3.8	110	0.00
9	Benzaldehyde	1.082	0.923	14.7	75	0.00
10 C	Phenol	1.620	1.555	4.0	109	0.00
11	bis(2-Chloroethyl)ether	1.263	1.211	4.1	109	0.00
12	1,3-Dichlorobenzene	1.499	1.414	5.7	108	0.00
13 C	1,4-Dichlorobenzene	1.532	1.451	5.3	109	0.00
14	1,2-Dichlorobenzene	1.483	1.402	5.5	108	0.00
15	Benzyl Alcohol	1.227	1.197	2.4	111	0.00
16	2,2'-oxybis(1-Chloropropane	1.692	1.619	4.3	110	0.00
17	2-Methylphenol	1.125	1.094	2.8	109	0.00
18	Hexachloroethane	0.563	0.536	4.8	109	0.00
19 P	n-Nitroso-di-n-propylamine	1.023	0.979	4.3	109	0.00
20	3+4-Methylphenols	1.560	1.506	3.5	109	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
22	Acetophenone	0.502	0.479	4.6	112	0.00
23 S	Nitrobenzene-d5	0.395	0.328	17.0	97	0.00
24	Nitrobenzene	0.353	0.330	6.5	108	0.00
25	Isophorone	0.700	0.662	5.4	111	0.00
26 C	2-Nitrophenol	0.181	0.176	2.8	110	0.00
27	2,4-Dimethylphenol	0.312	0.303	2.9	111	0.00
28	bis(2-Chloroethoxy)methane	0.423	0.400	5.4	111	0.00
29 C	2,4-Dichlorophenol	0.310	0.299	3.5	111	0.00
30	1,2,4-Trichlorobenzene	0.340	0.316	7.1	109	0.00
31	Naphthalene	1.040	0.973	6.4	110	0.00
32	Benzoic acid	0.230	0.220	4.3	110	0.01
33	4-Chloroaniline	0.459	0.437	4.8	109	0.00
34 C	Hexachlorobutadiene	0.211	0.196	7.1	108	0.00
35	Caprolactam	0.114	0.109	4.4	112	0.00
36 C	4-Chloro-3-methylphenol	0.355	0.345	2.8	113	0.00
37	2-Methylnaphthalene	0.676	0.640	5.3	111	0.00
38	1-Methylnaphthalene	0.719	0.675	6.1	111	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	0.578	0.547	5.4	111	0.00
41 P	Hexachlorocyclopentadiene	0.349	0.350	-0.3	115	0.00
42 S	2,4,6-Tribromophenol	0.269	0.241	10.4	105	0.00
43 C	2,4,6-Trichlorophenol	0.390	0.380	2.6	113	0.00
44	2,4,5-Trichlorophenol	0.427	0.418	2.1	112	0.00
45 S	2-Fluorobiphenyl	1.463	1.208	17.4	99	0.00
46	1,1'-Biphenyl	1.453	1.368	5.8	111	0.00
47	2-Chloronaphthalene	1.131	1.065	5.8	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081425\
 Data File : BP025400.D
 Acq On : 14 Aug 2025 18:05
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP081425

Quant Time: Aug 14 18:24:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.322	2.1	111	0.00
49	Acenaphthylene	1.871	1.764	5.7	111	0.00
50	Dimethylphthalate	1.465	1.398	4.6	114	0.00
51	2,6-Dinitrotoluene	0.309	0.303	1.9	112	0.00
52 C	Acenaphthene	1.098	1.011	7.9	111	0.00
53	3-Nitroaniline	0.347	0.348	-0.3	116	0.00
54 P	2,4-Dinitrophenol	0.165	0.172	-4.2	120	0.00
55	Dibenzofuran	1.736	1.620	6.7	110	0.00
56 P	4-Nitrophenol	0.285	0.283	0.7	115	0.00
57	2,4-Dinitrotoluene	0.440	0.445	-1.1	116	0.00
58	Fluorene	1.370	1.267	7.5	111	0.01
59	2,3,4,6-Tetrachlorophenol	0.377	0.369	2.1	115	0.00
60	Diethylphthalate	1.488	1.407	5.4	112	0.00
61	4-Chlorophenyl-phenylether	0.681	0.623	8.5	109	0.01
62	4-Nitroaniline	0.356	0.345	3.1	112	0.00
63	Azobenzene	1.274	1.227	3.7	113	0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.01
65	4,6-Dinitro-2-methylphenol	0.125	0.124	0.8	118	0.00
66 c	n-Nitrosodiphenylamine	0.610	0.568	6.9	112	0.00
67	4-Bromophenyl-phenylether	0.220	0.204	7.3	112	0.00
68	Hexachlorobenzene	0.255	0.236	7.5	114	0.01
69	Atrazine	0.228	0.217	4.8	116	0.00
70 C	Pentachlorophenol	0.161	0.156	3.1	116	0.01
71	Phenanthrene	1.099	1.001	8.9	113	0.01
72	Anthracene	1.122	1.054	6.1	114	0.00
73	Carbazole	1.057	0.981	7.2	114	0.00
74	Di-n-butylphthalate	1.300	1.222	6.0	115	0.02
75 C	Fluoranthene	1.311	1.228	6.3	117	0.02
76 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
77	Benzidine	0.683	0.721	-5.6	88	0.00
78	Pyrene	1.300	1.216	6.5	113	0.01
79 S	Terphenyl-d14	1.093	0.907	17.0	102	0.02
80	Butylbenzylphthalate	0.589	0.566	3.9	115	0.01
81	Benzo(a)anthracene	1.319	1.218	7.7	113	0.00
82	3,3'-Dichlorobenzidine	0.497	0.473	4.8	114	0.01
83	Chrysene	1.235	1.153	6.6	115	0.00
84	Bis(2-ethylhexyl)phthalate	0.867	0.840	3.1	119	0.00
85 c	Di-n-octyl phthalate	1.492	1.414	5.2	115	0.00
86 I	Perylene-d12	1.000	1.000	0.0	117	0.00
87	Indeno(1,2,3-cd)pyrene	1.467	1.361	7.2	113	0.01
88	Benzo(b)fluoranthene	1.179	1.090	7.5	113	0.02
89	Benzo(k)fluoranthene	1.180	1.105	6.4	115	0.01
90 C	Benzo(a)pyrene	1.148	1.066	7.1	113	0.02
91	Dibenzo(a,h)anthracene	1.199	1.123	6.3	114	0.00
92	Benzo(g,h,i)perylene	1.178	1.096	7.0	113	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081425\
Data File : BP025400.D
Acq On : 14 Aug 2025 18:05
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP081425

Quant Time: Aug 14 18:24:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 14 18:14:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0