

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2928

Client ID : SSTDCCC040

Date Analyzed: 08/27/2025

Lab File ID: BP025596.D

Time Analyzed: 09:39

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	161611	7.778	681785	10.54	463868	14.40
UPPER LIMIT	323222	8.278	1363570	11.043	927736	14.896
LOWER LIMIT	80805.5	7.278	340893	10.043	231934	13.896
EPA SAMPLE NO.						
01 PB169370BL	142863	7.77	561480	10.55	364523	14.40
02 PB169370BS	148040	7.78	615021	10.55	404292	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.