

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081125\
 Data File : VW032055.D
 Acq On : 11 Aug 2025 10:09
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/12/2025
 Supervised By :Semsettin Yesilyurt 08/12/2025

Quant Time: Aug 12 03:38:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 03:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	223324	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	406973	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	378090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	178995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	159972	49.529	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery =	99.060%		
35) Dibromofluoromethane	7.898	113	132195	50.279	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery =	100.560%		
50) Toluene-d8	10.325	98	506634	52.098	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery =	104.200%		
62) 4-Bromofluorobenzene	12.617	95	187331	52.274	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery =	104.540%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.040	85	75096	47.015	ug/l	100
3) Chloromethane	2.259	50	97719	45.418	ug/l	100
4) Vinyl Chloride	2.405	62	118378	45.486	ug/l	100
5) Bromomethane	2.820	94	91749	46.991	ug/l	100
6) Chloroethane	2.972	64	82045	47.641	ug/l	100
7) Trichlorofluoromethane	3.302	101	103706	49.743	ug/l	100
8) Diethyl Ether	3.722	74	84764	49.624	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.094	101	115526	46.968	ug/l	100
10) Methyl Iodide	4.307	142	167079	47.909	ug/l	100
11) Tert butyl alcohol	5.222	59	56322	269.294	ug/l	100
12) 1,1-Dichloroethene	4.082	96	128304	47.373	ug/l	100
13) Acrolein	3.930	56	78858	236.637	ug/l	100
14) Allyl chloride	4.704	41	192434	48.819	ug/l	100
15) Acrylonitrile	5.405	53	213941	262.963	ug/l	100
16) Acetone	4.161	43	182559	244.127	ug/l	100
17) Carbon Disulfide	4.423	76	347752	47.836	ug/l	100
18) Methyl Acetate	4.710	43	123902	52.361	ug/l	100
19) Methyl tert-butyl Ether	5.466	73	246045	51.560	ug/l	100
20) Methylene Chloride	4.960	84	152186	44.016	ug/l	100
21) trans-1,2-Dichloroethene	5.460	96	139994	48.707	ug/l	100
22) Diisopropyl ether	6.344	45	421731	50.751	ug/l	100
23) Vinyl Acetate	6.283	43	1450553	265.873	ug/l	100
24) 1,1-Dichloroethane	6.240	63	258279	48.646	ug/l	100
25) 2-Butanone	7.191	43	284711	267.883	ug/l	100
26) 2,2-Dichloropropane	7.191	77	129390	48.578	ug/l	100
27) cis-1,2-Dichloroethene	7.191	96	167240	49.779	ug/l	100
28) Bromochloromethane	7.533	49	121266	50.191	ug/l	100
29) Tetrahydrofuran	7.551	42	190719	272.100	ug/l	100
30) Chloroform	7.691	83	275905	49.018	ug/l	100
31) Cyclohexane	7.971	56	214597	46.622	ug/l	100
32) 1,1,1-Trichloroethane	7.886	97	194676	47.617	ug/l	100
36) 1,1-Dichloropropene	8.099	75	190046	51.274	ug/l	100
37) Ethyl Acetate	7.276	43	119689	54.829	ug/l	100
38) Carbon Tetrachloride	8.087	117	182422	51.000	ug/l	100
39) Methylcyclohexane	9.343	83	235341	52.093	ug/l	100
40) Benzene	8.337	78	597760	51.883	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081125\
 Data File : VW032055.D
 Acq On : 11 Aug 2025 10:09
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/12/2025
 Supervised By :Semsettin Yesilyurt 08/12/2025

Quant Time: Aug 12 03:38:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 03:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.496	41	68136	55.039	ug/l	100
42) 1,2-Dichloroethane	8.410	62	189755	51.471	ug/l	100
43) Isopropyl Acetate	8.435	43	219064	54.544	ug/l	100
44) Trichloroethene	9.099	130	137910	50.264	ug/l	100
45) 1,2-Dichloropropane	9.374	63	143446	51.537	ug/l	100
46) Dibromomethane	9.465	93	93001	51.770	ug/l	100
47) Bromodichloromethane	9.654	83	216521	52.307	ug/l	100
48) Methyl methacrylate	9.441	41	108287	55.849	ug/l	100
49) 1,4-Dioxane	9.465	88	26579	1144.321	ug/l	100
51) 4-Methyl-2-Pentanone	10.215	43	635174	280.194	ug/l	100
52) Toluene	10.392	92	379230	52.258	ug/l	100
53) t-1,3-Dichloropropene	10.611	75	209555	54.035	ug/l	100
54) cis-1,3-Dichloropropene	10.075	75	236454	53.744	ug/l	100
55) 1,1,2-Trichloroethane	10.788	97	124341	52.088	ug/l	100
56) Ethyl methacrylate	10.648	69	178974	56.027	ug/l	100
57) 1,3-Dichloropropane	10.934	76	216719	52.084	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.928	63	490522	277.613	ug/l	100
59) 2-Hexanone	10.971	43	447423	284.986	ug/l	100
60) Dibromochloromethane	11.129	129	142234	52.161	ug/l	100
61) 1,2-Dibromoethane	11.233	107	121625	52.430	ug/l	100
64) Tetrachloroethene	10.867	164	105400	48.263	ug/l	100
65) Chlorobenzene	11.654	112	403926	49.082	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.727	131	128215	49.935	ug/l	100
67) Ethyl Benzene	11.727	91	706877	50.709	ug/l	100
68) m/p-Xylenes	11.837	106	552442	102.704	ug/l	100
69) o-Xylene	12.166	106	257591	51.812	ug/l	100
70) Styrene	12.178	104	457949	51.858	ug/l	100
71) Bromoform	12.349	173	77959	53.078	ug/l #	100
73) Isopropylbenzene	12.465	105	654522	52.323	ug/l	100
74) N-amyl acetate	12.269	43	199256	55.520	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.708	83	156076	52.549	ug/l	100
76) 1,2,3-Trichloropropane	12.763	75	137632m	50.019	ug/l	
77) Bromobenzene	12.745	156	156335	52.326	ug/l	100
78) n-propylbenzene	12.800	91	805515	52.304	ug/l	100
79) 2-Chlorotoluene	12.891	91	490665	51.622	ug/l	100
80) 1,3,5-Trimethylbenzene	12.940	105	555481	52.305	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.507	75	51811	55.213	ug/l	100
82) 4-Chlorotoluene	12.983	91	519957	51.488	ug/l	100
83) tert-Butylbenzene	13.202	119	462373	52.460	ug/l	100
84) 1,2,4-Trimethylbenzene	13.245	105	572429	52.950	ug/l	100
85) sec-Butylbenzene	13.379	105	690647	51.829	ug/l	100
86) p-Isopropyltoluene	13.495	119	589769	53.089	ug/l	100
87) 1,3-Dichlorobenzene	13.495	146	313624	52.754	ug/l	100
88) 1,4-Dichlorobenzene	13.574	146	299453	49.844	ug/l	100
89) n-Butylbenzene	13.818	91	569364	52.756	ug/l	100
90) Hexachloroethane	14.086	117	101464	49.942	ug/l	100
91) 1,2-Dichlorobenzene	13.867	146	273581	50.801	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.476	75	28437	53.857	ug/l	100
93) 1,2,4-Trichlorobenzene	15.129	180	164726	51.132	ug/l	100
94) Hexachlorobutadiene	15.226	225	68983	47.902	ug/l	100
95) Naphthalene	15.354	128	434944	55.099	ug/l	100
96) 1,2,3-Trichlorobenzene	15.549	180	149556	51.022	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081125\
Data File : VW032055.D
Acq On : 11 Aug 2025 10:09
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/12/2025
Supervised By :Semsettin Yesilyurt 08/12/2025

Quant Time: Aug 12 03:38:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W081125S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 03:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

