

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: HERO01
 Lab Code: ACE SDG NO.: Q2928
 Lab File ID: VW032107.D Date Analyzed: 08/22/2025
 Instrument ID: MSVOA_W Time Analyzed: 12:12
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	229263	7.96	385827	8.85	350083	11.63
UPPER LIMIT	458526	8.459	771654	9.349	700166	12.129
LOWER LIMIT	114632	7.459	192914	8.349	175042	11.129
EPA SAMPLE NO.						
27-BELL	183461	7.97	365151	8.85	344552	11.64
VW0822SBL01	152490	7.96	320732	8.85	308340	11.64
VW0822SBS01	205288	7.95	359763	8.85	335789	11.63
VW0822SBSD01	208637	7.97	348689	8.85	334530	11.64

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.