

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: HERO01  
 Lab Code: ACE SDG NO.: Q2928  
 Lab File ID: VW032107.D Date Analyzed: 08/22/2025  
 Instrument ID: MSVOA\_W Time Analyzed: 12:12  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 177844        | 13.556 |  |  |  |  |
| UPPER LIMIT    | 355688        | 14.056 |  |  |  |  |
| LOWER LIMIT    | 88922         | 13.056 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| 27-BELL        | 168107        | 13.56  |  |  |  |  |
| VW0822SBL01    | 143329        | 13.56  |  |  |  |  |
| VW0822SBS01    | 176918        | 13.56  |  |  |  |  |
| VW0822SBSD01   | 180581        | 13.56  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.