

### Initial and Continuing Calibration Blank Summary

**Client:** ATG-GREENVILLE AEC

**SDG No.:** Q2929

**Project:** AEC-2025-0013- 500 Sterling Ave - Schenectady NY

**RunNo.:** LB136918

| Analyte                       | Units | Result   | Acceptance<br>Limits | Conc<br>Qual | MDL  | RDL | Analysis<br>Date |
|-------------------------------|-------|----------|----------------------|--------------|------|-----|------------------|
| Sample ID: <b>ICB</b><br>COD  | mg/L  | < 5.0000 | 5.0000               | U            | 1.50 | 10  | 05/28/2025       |
| Sample ID: <b>CCB1</b><br>COD | mg/L  | < 5.0000 | 5.0000               | U            | 1.50 | 10  | 08/22/2025       |
| Sample ID: <b>CCB2</b><br>COD | mg/L  | < 5.0000 | 5.0000               | U            | 1.50 | 10  | 08/22/2025       |
| Sample ID: <b>CCB3</b><br>COD | mg/L  | < 5.0000 | 5.0000               | U            | 1.50 | 10  | 08/22/2025       |