

Surrogate Summary

SDG No.: Q2929

Client: ATG-GREENVILLE AEC

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2929-01	DA3-SW2	1,2-Dichloroethane-d4	50	50.6	101		74	125
		Dibromofluoromethane	50	40.5	81		75	124
		Toluene-d8	50	44.3	89		86	113
		4-Bromofluorobenzene	50	41.3	83		77	121
Q2929-02	DA5-SW3	1,2-Dichloroethane-d4	50	49.5	99		74	125
		Dibromofluoromethane	50	43.0	86		75	124
		Toluene-d8	50	45.6	91		86	113
		4-Bromofluorobenzene	50	50.7	101		77	121
Q2929-03	DA5-SW4	1,2-Dichloroethane-d4	50	59.5	119		74	125
		Dibromofluoromethane	50	45.4	91		75	124
		Toluene-d8	50	45.9	92		86	113
		4-Bromofluorobenzene	50	46.2	92		77	121
Q2929-04	DA6-SW5	1,2-Dichloroethane-d4	50	54.2	108		74	125
		Dibromofluoromethane	50	46.6	93		75	124
		Toluene-d8	50	44.7	89		86	113
		4-Bromofluorobenzene	50	38.9	78		77	121
VV0821WBL01	VV0821WBL01	1,2-Dichloroethane-d4	50	48.2	96		74	125
		Dibromofluoromethane	50	43.8	88		75	124
		Toluene-d8	50	43.4	87		86	113
		4-Bromofluorobenzene	50	42.3	85		77	121
VV0821WBS01	VV0821WBS01	1,2-Dichloroethane-d4	50	55.0	110		74	125
		Dibromofluoromethane	50	44.8	90		75	124
		Toluene-d8	50	49.6	99		86	113
		4-Bromofluorobenzene	50	48.8	98		77	121
VV0821WBSD0	VV0821WBSD01	1,2-Dichloroethane-d4	50	51.2	102		74	125
		Dibromofluoromethane	50	49.2	98		75	124
		Toluene-d8	50	48.9	98		86	113
		4-Bromofluorobenzene	50	49.0	98		77	121
VV0822WBL01	VV0822WBL01	1,2-Dichloroethane-d4	50	60.0	120		74	125
		Dibromofluoromethane	50	47.4	95		75	124
		Toluene-d8	50	47.7	95		86	113
		4-Bromofluorobenzene	50	49.3	99		77	121
VV0822WBS01	VV0822WBS01	1,2-Dichloroethane-d4	50	48.5	97		74	125
		Dibromofluoromethane	50	46.0	92		75	124
		Toluene-d8	50	46.3	93		86	113
		4-Bromofluorobenzene	50	48.5	97		77	121