

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2929</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ATG-GREENVILLE AEC</u>	Datafile :	<u>VV038986.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VV0821WBSD01	Benzene	20	20.0	ug/L	100	4		82	109	15
	Toluene	20	20.3	ug/L	102	5		82	110	16
	Ethyl Benzene	20	19.9	ug/L	100	4		83	109	16
	m/p-Xylenes	40	41.8	ug/L	104	13		82	110	15
	o-Xylene	20	20.5	ug/L	103	5		83	109	20