

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2929</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ATG-GREENVILLE AEC</u>	Datafile :	<u>VV039011.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VV0822WBS01	Benzene	20	19.4	ug/L	97			82	109	
	Toluene	20	19.6	ug/L	98			82	110	
	Ethyl Benzene	20	20.7	ug/L	104			83	109	
	m/p-Xylenes	40	41.0	ug/L	103			82	110	
	o-Xylene	20	20.8	ug/L	104			83	109	