

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2930
Project:	1-4 Industry Drive - Waterford NY	Sample ID:	Q2931-01
Client ID:	DA1-1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	55.0		13.4		50.0	1	83		08/22/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2930
Project:	1-4 Industry Drive - Waterford NY	Sample ID:	Q2931-01
Client ID:	DA1-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	52.0		13.4		50.0	1	77		08/22/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2930
Project:	1-4 Industry Drive - Waterford NY	Sample ID:	Q2935-01
Client ID:	EFFLUENTMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	35.2		15.1		20.0	1	101		08/25/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2930
Project:	1-4 Industry Drive - Waterford NY	Sample ID:	Q2935-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	35.5		15.1		20.0	1	102		08/25/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2930
Project:	1-4 Industry Drive - Waterford NY	Sample ID:	Q2942-01
Client ID:	MH-8-22-25MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	70.5		50.5		20.0	1	100		08/25/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2930
Project:	1-4 Industry Drive - Waterford NY	Sample ID:	Q2942-01
Client ID:	MH-8-22-25MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	70.8		50.5		20.0	1	102		08/25/2025