

### Duplicate Sample Summary

|                   |                                   |                                         |          |
|-------------------|-----------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ATG-GREENVILLE AEC                | <b>SDG No.:</b>                         | Q2931    |
| <b>Project:</b>   | AEC-2025-0013-CSC 4031 Watervliet | <b>Sample ID:</b>                       | Q2931-01 |
| <b>Client ID:</b> | DA1-1DUP                          | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| COD     | mg/L  | +/-20               | 13.4             |                    | 13.4                |                    | 1                  | 0          |      | 08/22/2025       |

### Duplicate Sample Summary

|                   |                                   |                                         |          |
|-------------------|-----------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ATG-GREENVILLE AEC                | <b>SDG No.:</b>                         | Q2931    |
| <b>Project:</b>   | AEC-2025-0013-CSC 4031 Watervliet | <b>Sample ID:</b>                       | Q2931-01 |
| <b>Client ID:</b> | DA1-1MSD                          | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| COD     | mg/L  | +/-20               | 55.0             |                    | 52.0                |                    | 1                  | 5.61       |      | 08/22/2025       |

### Duplicate Sample Summary

|                   |                                   |                                         |          |
|-------------------|-----------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ATG-GREENVILLE AEC                | <b>SDG No.:</b>                         | Q2931    |
| <b>Project:</b>   | AEC-2025-0013-CSC 4031 Watervliet | <b>Sample ID:</b>                       | Q2935-01 |
| <b>Client ID:</b> | EFFLUENTMSD                       | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte        | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|----------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Oil and Grease | mg/L  | +/-18               | 35.2             |                    | 35.5                |                    | 1                  | 0.85       |      | 08/25/2025       |

## Duplicate Sample Summary

|                                                   |                                           |
|---------------------------------------------------|-------------------------------------------|
| <b>Client:</b> ATG-GREENVILLE AEC                 | <b>SDG No.:</b> Q2931                     |
| <b>Project:</b> AEC-2025-0013-CSC 4031 Watervliet | <b>Sample ID:</b> Q2942-01                |
| <b>Client ID:</b> MH-8-22-25MSD                   | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte        | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|----------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Oil and Grease | mg/L  | +/-18               | 70.5             |                    | 70.8                |                    | 1                  | 0.42       |      | 08/25/2025       |