

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q2931

Project: AEC-2025-0013-CSC 4031 Watervliet

Run No.: LB136918

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136918BS							
COD	mg/L	50	49.9		100	1	90-110	08/22/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q2931

Project: AEC-2025-0013-CSC 4031 Watervliet

Run No.: LB136941

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136941BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	08/25/2025