

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2931

Analytical Method: SW8260-Low

Client: ATG-GREENVILLE AEC

Datafile : VV038985.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VV0821WBS01	Benzene	20	19.2	ug/L	96			82	109	
	Toluene	20	19.4	ug/L	97			82	110	
	Ethyl Benzene	20	19.1	ug/L	96			83	109	
	m/p-Xylenes	40	36.4	ug/L	91			82	110	
	o-Xylene	20	19.6	ug/L	98			83	109	