

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ATGG01
 Lab Code: ACE SDG NO.: Q2931
 Lab File ID: VV038982.D Date Analyzed: 08/21/2025
 Instrument ID: MSVOA_V Time Analyzed: 15:42
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	531490	4.60	833989	5.53	788137	8.77
UPPER LIMIT	1062980	5.096	1667980	6.029	1576270	9.273
LOWER LIMIT	265745	4.096	416995	5.029	394069	8.273
EPA SAMPLE NO.						
DA1-1	547704	4.60	987111	5.53	888922	8.78
VV0821WBL01	502511	4.60	836910	5.53	730493	8.78
VV0821WBS01	490499	4.60	863998	5.53	809489	8.77
VV0821WBSD01	484688	4.60	811429	5.53	760113	8.78

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.