

### Laboratory Control Sample Summary

**Client:** Portal Partners Tri-Venture

**SDG No.:** Q2934

**Project:** Amtrak Sawtooth Bridges 2025

**Run No.:** LB136915

| Analyte             | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|---------------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID           | lb136915BS |            |        |                 |            |                 |                     |               |
| Hexavalent Chromium | mg/L       | 0.5        | 0.51   |                 | 102        | 1               | 90-111              | 08/21/2025    |