

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 08/22/2025 Time : 10:30 Temp : 97 °C

End Digest Date: 08/22/2025 Time : 13:32 Temp : 98 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: SPS

Supervisor Signature: [Signature]

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6180
LFS-2	0.25	M6181
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP85156
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 97C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
08/22/25 14:32	SPS met dia.	[Signature] Metals Lab
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB169358BL	PBW358	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	1
PB169358BS	LCS358	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	2
Q2927-01	A6121M	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
Q2934-01	ENV-SW01	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
Q2934-02	ENV-SW02	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	5
Q2934-02MS	ENV-SW02MS	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	7
Q2934-02MSD	ENV-SW02MSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	8
Q2934-02DUP	ENV-SW02DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	6
Q2934-03	ENV-SW03	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	9
Q2934-04	ENV-SW04	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	10