

Continuing Calibration Report for SequenceID : FD082625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049670.D

Aromatic C10-C12	4267425.000	40.000	4.405	6.149	106685.625	107939.663	1.162
Aromatic C12-C16	7383707.000	60.000	6.150	8.777	123061.783	123580.152	0.419
Aromatic C16-C21	9769945.000	80.000	8.778	13.069	122124.313	126486.361	3.449
Aromatic C21-C36	20299067.000	180.000	13.070	18.490	112772.594	112618.254	-0.137
Aromatic EPH	41720144.000	360.000	4.405	18.490	115889.289	117007.195	0.955