

Continuing Calibration Report for SequenceID : FD082625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049681.D

Aromatic C10-C12	4425555.000	40.000	4.405	6.149	110638.875	107939.663	-2.501
Aromatic C12-C16	7638456.000	60.000	6.150	8.777	127307.600	123580.152	-3.016
Aromatic C16-C21	10201141.000	80.000	8.778	13.069	127514.263	126486.361	-0.813
Aromatic C21-C36	21291076.000	180.000	13.070	18.490	118283.756	112618.254	-5.031
Aromatic EPH	43556228.000	360.000	4.405	18.490	120989.522	117007.195	-3.403