

Continuing Calibration Report for SequenceID : FD082725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049684.D

Aromatic C10-C12	4799858.000	40.000	4.406	6.149	119996.450	107939.663	-11.170
Aromatic C12-C16	8371051.000	60.000	6.150	8.778	139517.517	123580.152	-12.896
Aromatic C16-C21	10864762.000	80.000	8.779	13.069	135809.525	126486.361	-7.371
Aromatic C21-C36	21769018.000	180.000	13.070	18.489	120938.989	112618.254	-7.388
Aromatic EPH	45804689.000	360.000	4.406	18.489	127235.247	117007.195	-8.741