

Continuing Calibration Report for SequenceID : FD082725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049687.D

Aromatic C10-C12	4723879.000	40.000	4.406	6.149	118096.975	107939.663	-9.410
Aromatic C12-C16	8202684.000	60.000	6.150	8.778	136711.400	123580.152	-10.626
Aromatic C16-C21	10706166.000	80.000	8.779	13.069	133827.075	126486.361	-5.804
Aromatic C21-C36	21685094.000	180.000	13.070	18.489	120472.744	112618.254	-6.974
Aromatic EPH	45317823.000	360.000	4.406	18.489	125882.842	117007.195	-7.586