

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2934

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 15:40

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.55       | 5.55      | 5.45      | 5.65 | 0.00       |
| Aroclor-1254-2       | (2) | 5.70       | 5.70      | 5.60      | 5.80 | 0.00       |
| Aroclor-1254-3       | (3) | 6.10       | 6.11      | 6.01      | 6.21 | 0.01       |
| Aroclor-1254-4       | (4) | 6.33       | 6.34      | 6.24      | 6.44 | 0.01       |
| Aroclor-1254-5       | (5) | 6.75       | 6.75      | 6.65      | 6.85 | 0.00       |
| Tetrachloro-m-xylene |     | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   |     | 8.68       | 8.68      | 8.58      | 8.78 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2934

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 15:40

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.52       | 5.53      | 5.43      | 5.63 | 0.01       |
| Aroclor-1254-2 (2)   | 5.67       | 5.67      | 5.57      | 5.77 | 0.00       |
| Aroclor-1254-3 (3)   | 6.07       | 6.07      | 5.97      | 6.17 | 0.00       |
| Aroclor-1254-4 (4)   | 6.30       | 6.30      | 6.20      | 6.40 | 0.00       |
| Aroclor-1254-5 (5)   | 6.72       | 6.72      | 6.62      | 6.82 | 0.00       |
| Tetrachloro-m-xylene | 3.65       | 3.66      | 3.56      | 3.76 | 0.01       |
| Decachlorobiphenyl   | 8.63       | 8.63      | 8.53      | 8.73 | 0.00       |