

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2934

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/19/2025 08/19/2025

Continuing Calib Time: 14:14

Initial Calibration Time(s): 10:02 17:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.90	5.89	5.79	5.99	-0.01
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.29	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.66	7.65	7.55	7.75	-0.01
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.57	8.56	8.46	8.66	-0.01
Tetrachloro-m-xylene		4.66	4.65	4.55	4.75	-0.01
Decachlorobiphenyl		10.44	10.43	10.33	10.53	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2934

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/19/2025 08/19/2025

Continuing Calib Time: 14:14

Initial Calibration Time(s): 10:02 17:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-2	(2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-3	(3)	5.08	5.08	4.98	5.18	0.00
Aroclor-1016-4	(4)	5.12	5.12	5.02	5.22	0.00
Aroclor-1016-5	(5)	5.33	5.33	5.23	5.43	0.00
Aroclor-1260-1	(1)	6.36	6.36	6.26	6.46	0.00
Aroclor-1260-2	(2)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-3	(3)	6.71	6.70	6.60	6.80	0.00
Aroclor-1260-4	(4)	7.17	7.17	7.07	7.27	0.00
Aroclor-1260-5	(5)	7.41	7.41	7.31	7.51	0.00
Tetrachloro-m-xylene		3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl		8.82	8.82	8.72	8.92	0.00