

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2934
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113208.D **Time Analyzed:** 02:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	520.600	500.000	4.1
Aroclor-1016-2	4.765	4.666	4.866	533.990	500.000	6.8
Aroclor-1016-3	4.822	4.724	4.924	533.000	500.000	6.6
Aroclor-1016-4	4.941	4.843	5.043	481.610	500.000	-3.7
Aroclor-1016-5	5.198	5.099	5.299	574.940	500.000	15.0
Aroclor-1260-1	6.233	6.136	6.336	517.130	500.000	3.4
Aroclor-1260-2	6.423	6.326	6.526	525.310	500.000	5.1
Aroclor-1260-3	6.788	6.691	6.891	521.760	500.000	4.4
Aroclor-1260-4	7.047	6.951	7.151	526.560	500.000	5.3
Aroclor-1260-5	7.291	7.194	7.394	520.660	500.000	4.1
Decachlorobiphenyl	8.679	8.582	8.782	48.780	50.000	-2.4
Tetrachloro-m-xylene	3.662	3.563	3.763	52.400	50.000	4.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2934
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113208.D **Time Analyzed:** 02:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.727	4.629	4.829	501.500	500.000	0.3
Aroclor-1016-2	4.745	4.647	4.847	494.870	500.000	-1.0
Aroclor-1016-3	4.920	4.822	5.022	486.480	500.000	-2.7
Aroclor-1016-4	4.963	4.864	5.064	471.280	500.000	-5.7
Aroclor-1016-5	5.175	5.076	5.276	526.100	500.000	5.2
Aroclor-1260-1	6.202	6.104	6.304	474.200	500.000	-5.2
Aroclor-1260-2	6.390	6.292	6.492	477.900	500.000	-4.4
Aroclor-1260-3	6.541	6.443	6.643	441.190	500.000	-11.8
Aroclor-1260-4	7.010	6.912	7.112	455.950	500.000	-8.8
Aroclor-1260-5	7.253	7.155	7.355	445.530	500.000	-10.9
Decachlorobiphenyl	8.625	8.528	8.728	49.550	50.000	-0.9
Tetrachloro-m-xylene	3.654	3.555	3.755	50.210	50.000	0.4