

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2934
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074607.D **Time Analyzed:** 01:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.810	5.706	5.906	517.150	500.000	3.4
Aroclor-1016-2	5.832	5.727	5.927	535.080	500.000	7.0
Aroclor-1016-3	5.894	5.790	5.990	490.880	500.000	-1.8
Aroclor-1016-4	5.992	5.888	6.088	514.370	500.000	2.9
Aroclor-1016-5	6.284	6.180	6.380	532.650	500.000	6.5
Aroclor-1260-1	7.401	7.297	7.497	502.910	500.000	0.6
Aroclor-1260-2	7.653	7.550	7.750	497.210	500.000	-0.6
Aroclor-1260-3	8.011	7.908	8.108	526.350	500.000	5.3
Aroclor-1260-4	8.240	8.136	8.336	506.800	500.000	1.4
Aroclor-1260-5	8.565	8.462	8.662	489.770	500.000	-2.0
Decachlorobiphenyl	10.436	10.333	10.533	50.680	50.000	1.4
Tetrachloro-m-xylene	4.659	4.554	4.754	52.580	50.000	5.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2934
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074607.D **Time Analyzed:** 01:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.900	4.799	4.999	528.170	500.000	5.6
Aroclor-1016-2	4.958	4.855	5.055	550.600	500.000	10.1
Aroclor-1016-3	5.079	4.976	5.176	521.820	500.000	4.4
Aroclor-1016-4	5.119	5.017	5.217	501.960	500.000	0.4
Aroclor-1016-5	5.334	5.231	5.431	472.280	500.000	-5.5
Aroclor-1260-1	6.363	6.261	6.461	480.650	500.000	-3.9
Aroclor-1260-2	6.549	6.448	6.648	458.750	500.000	-8.3
Aroclor-1260-3	6.703	6.602	6.802	420.480	500.000	-15.9
Aroclor-1260-4	7.173	7.072	7.272	436.150	500.000	-12.8
Aroclor-1260-5	7.412	7.311	7.511	431.770	500.000	-13.6
Decachlorobiphenyl	8.814	8.715	8.915	42.100	50.000	-15.8
Tetrachloro-m-xylene	3.801	3.699	3.899	55.450	50.000	10.9