

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** PORT06  
**Lab Code:** ACE **SDG NO.:** Q2934  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL04 **Date Analyzed:** 08/23/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074619.D **Time Analyzed:** 05:26

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.806  | 5.706     | 5.906  | 477.270            | 500.000           | -4.5 |
| Aroclor-1016-2       | 5.828  | 5.727     | 5.927  | 504.260            | 500.000           | 0.9  |
| Aroclor-1016-3       | 5.891  | 5.790     | 5.990  | 467.730            | 500.000           | -6.5 |
| Aroclor-1016-4       | 5.987  | 5.888     | 6.088  | 482.490            | 500.000           | -3.5 |
| Aroclor-1016-5       | 6.280  | 6.180     | 6.380  | 500.490            | 500.000           | 0.1  |
| Aroclor-1260-1       | 7.397  | 7.297     | 7.497  | 472.640            | 500.000           | -5.5 |
| Aroclor-1260-2       | 7.650  | 7.550     | 7.750  | 463.510            | 500.000           | -7.3 |
| Aroclor-1260-3       | 8.008  | 7.908     | 8.108  | 491.980            | 500.000           | -1.6 |
| Aroclor-1260-4       | 8.236  | 8.136     | 8.336  | 464.640            | 500.000           | -7.1 |
| Aroclor-1260-5       | 8.561  | 8.462     | 8.662  | 454.250            | 500.000           | -9.2 |
| Decachlorobiphenyl   | 10.430 | 10.333    | 10.533 | 47.100             | 50.000            | -5.8 |
| Tetrachloro-m-xylene | 4.655  | 4.554     | 4.754  | 49.100             | 50.000            | -1.8 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** PORT06  
**Lab Code:** ACE **SDG NO.:** Q2934  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL04 **Date Analyzed:** 08/23/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074619.D **Time Analyzed:** 05:26

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.897 | 4.799     | 4.999 | 480.720            | 500.000           | -3.9  |
| Aroclor-1016-2       | 4.955 | 4.855     | 5.055 | 492.480            | 500.000           | -1.5  |
| Aroclor-1016-3       | 5.075 | 4.976     | 5.176 | 473.430            | 500.000           | -5.3  |
| Aroclor-1016-4       | 5.116 | 5.017     | 5.217 | 458.630            | 500.000           | -8.3  |
| Aroclor-1016-5       | 5.330 | 5.231     | 5.431 | 420.760            | 500.000           | -15.8 |
| Aroclor-1260-1       | 6.359 | 6.261     | 6.461 | 454.660            | 500.000           | -9.1  |
| Aroclor-1260-2       | 6.545 | 6.448     | 6.648 | 459.130            | 500.000           | -8.2  |
| Aroclor-1260-3       | 6.700 | 6.602     | 6.802 | 406.660            | 500.000           | -18.7 |
| Aroclor-1260-4       | 7.169 | 7.072     | 7.272 | 439.820            | 500.000           | -12.0 |
| Aroclor-1260-5       | 7.408 | 7.311     | 7.511 | 433.530            | 500.000           | -13.3 |
| Decachlorobiphenyl   | 8.810 | 8.715     | 8.915 | 40.970             | 50.000            | -18.1 |
| Tetrachloro-m-xylene | 3.798 | 3.699     | 3.899 | 53.390             | 50.000            | 6.8   |