

Surrogate Summary

SDG No.: Q2934

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113002.D	PIBLK-PO113002.D	Tetrachloro-m-xyl	1	20	16.6	83		70 (60)	130 (140)
		Decachlorobiphen	1	20	17.2	86		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	16.9	84		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.1	91		70 (60)	130 (140)
I.BLK-PO113197.D	PIBLK-PO113197.D	Tetrachloro-m-xyl	1	20	17.8	89		70 (60)	130 (140)
		Decachlorobiphen	1	20	16.3	82		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.3	86		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.2	91		70 (60)	130 (140)
Q2934-01	ENV-SW01	Tetrachloro-m-xyl	1	20	11.1	56		30 (30)	150 (173)
		Decachlorobiphen	1	20	7.65	38		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	11.1	56		30 (30)	150 (173)
		Decachlorobiphen	2	20	7.90	40		30 (10)	150 (173)
Q2934-02	ENV-SW02	Tetrachloro-m-xyl	1	20	11.0	55		30 (30)	150 (173)
		Decachlorobiphen	1	20	8.53	43		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	10.8	54		30 (30)	150 (173)
		Decachlorobiphen	2	20	8.93	45		30 (10)	150 (173)
I.BLK-PO113212.D	PIBLK-PO113212.D	Tetrachloro-m-xyl	1	20	17.7	88		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.2	91		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.2	86		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.6	88		70 (60)	130 (140)
I.BLK-PP074490.D	PIBLK-PP074490.D	Tetrachloro-m-xyl	1	20	17.8	89		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.2	91		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	15.7	78		70 (60)	130 (140)
		Decachlorobiphen	2	20	16.9	85		70 (60)	130 (140)
I.BLK-PP074611.D	PIBLK-PP074611.D	Tetrachloro-m-xyl	1	20	18.1	90		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.0	95		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	15.2	76		70 (60)	130 (140)
		Decachlorobiphen	2	20	13.8	69	*	70 (60)	130 (140)
PB169361BL	PB169361BL	Tetrachloro-m-xyl	1	20	22.1	110		30 (30)	150 (173)
		Decachlorobiphen	1	20	23.4	117		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	19.8	99		30 (30)	150 (173)
		Decachlorobiphen	2	20	17.1	86		30 (10)	150 (173)
PB169361BS	PB169361BS	Tetrachloro-m-xyl	1	20	22.7	114		30 (30)	150 (173)
		Decachlorobiphen	1	20	23.5	118		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	20.2	101		30 (30)	150 (173)
		Decachlorobiphen	2	20	17.2	86		30 (10)	150 (173)
PB169361BSD	PB169361BSD	Tetrachloro-m-xyl	1	20	22.4	112		30 (30)	150 (173)
		Decachlorobiphen	1	20	23.2	116		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	20.0	100		30 (30)	150 (173)
		Decachlorobiphen	2	20	17.0	85		30 (10)	150 (173)
Q2934-03	ENV-SW03	Tetrachloro-m-xyl	1	20	12.1	60		30 (30)	150 (173)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q2934

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2934-03	ENV-SW03	Decachlorobiphen	1	20	9.91	50		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	9.78	49		30 (30)	150 (173)
		Decachlorobiphen	2	20	6.69	33		30 (10)	150 (173)
Q2934-04	ENV-SW04	Tetrachloro-m-xyl	1	20	7.95	40		30 (30)	150 (173)
		Decachlorobiphen	1	20	7.04	35		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	5.94	30		30 (30)	150 (173)
		Decachlorobiphen	2	20	4.77	24	*	30 (10)	150 (173)
		Tetrachloro-m-xyl	1	20	18.7	94		70 (60)	130 (140)
I.BLK-PP074623.D	PIBLK-PP074623.D	Decachlorobiphen	1	20	19.6	98		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	15.3	77		70 (60)	130 (140)
		Decachlorobiphen	2	20	14.2	71		70 (60)	130 (140)