

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2934</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Portal Partners Tri-Venture</u>	Datafile :	<u>PP074614.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169361BSD (Column 1)	AR1016	5	4.50	ug/L	90	0			40 (77)	140 (107)	20 (20)
	AR1260	5	4.60	ug/L	92	7			40 (66)	140 (113)	20 (20)
PB169361BSD (Column 2)	AR1016	5	4.00	ug/L	80	3			40 (77)	140 (107)	20 (20)
	AR1260	5	4.30	ug/L	86	15			40 (66)	140 (113)	20 (20)