

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PORT06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2934</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/22/2025 10:49</u>
Lab File ID:	<u>BP025544.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.204	1.201		-0.2	
Benzaldehyde	1.082	1.377		27.3	
Phenol-d6	1.544	1.586		2.7	
Phenol	1.620	1.656		2.2	20.0
bis(2-Chloroethyl)ether	1.263	1.266		0.2	
2-Chlorophenol	1.359	1.376		1.3	
2-Methylphenol	1.125	1.147		2.0	
2,2-oxybis(1-Chloropropane)	1.692	1.738		2.7	
Acetophenone	0.502	0.494		-1.6	
3+4-Methylphenols	1.560	1.584		1.5	
n-Nitroso-di-n-propylamine	1.023	1.030	0.050	0.7	
Nitrobenzene-d5	0.395	0.396		0.3	
Hexachloroethane	0.563	0.554		-1.6	
Nitrobenzene	0.353	0.357		1.1	
Isophorone	0.700	0.697		-0.4	
2-Nitrophenol	0.181	0.184		1.7	20.0
2,4-Dimethylphenol	0.312	0.307		-1.6	
bis(2-Chloroethoxy)methane	0.423	0.416		-1.7	
2,4-Dichlorophenol	0.310	0.307		-1.0	20.0
Naphthalene	1.040	1.006		-3.3	
4-Chloroaniline	0.459	0.460		0.2	
Hexachlorobutadiene	0.211	0.200		-5.2	20.0
Caprolactam	0.114	0.124		8.8	
4-Chloro-3-methylphenol	0.355	0.364		2.5	20.0
2-Methylnaphthalene	0.676	0.652		-3.5	
Hexachlorocyclopentadiene	0.349	0.397	0.050	13.8	
2,4,6-Trichlorophenol	0.390	0.387		-0.8	20.0
2-Fluorobiphenyl	1.463	1.364		-6.8	
2,4,5-Trichlorophenol	0.427	0.430		0.7	
1,1-Biphenyl	1.453	1.392		-4.2	
2-Chloronaphthalene	1.131	1.080		-4.5	
2-Nitroaniline	0.329	0.351		6.7	
Dimethylphthalate	1.465	1.455		-0.7	
Acenaphthylene	1.871	1.834		-2.0	
2,6-Dinitrotoluene	0.309	0.324		4.9	
3-Nitroaniline	0.347	0.370		6.6	
Acenaphthene	1.098	1.042		-5.1	20.0
2,4-Dinitrophenol	0.165	0.228	0.050	38.2	
4-Nitrophenol	0.285	0.311	0.050	9.1	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PORT06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2934</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/22/2025 10:49</u>
Lab File ID:	<u>BP025544.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.736	1.669		-3.9	
2,4-Dinitrotoluene	0.440	0.471		7.0	
Diethylphthalate	1.488	1.520		2.2	
4-Chlorophenyl-phenylether	0.681	0.653		-4.1	
Fluorene	1.370	1.350		-1.5	
4-Nitroaniline	0.356	0.382		7.3	
4,6-Dinitro-2-methylphenol	0.125	0.146		16.8	
n-Nitrosodiphenylamine	0.610	0.590		-3.3	20.0
2,4,6-Tribromophenol	0.269	0.273		1.5	
4-Bromophenyl-phenylether	0.220	0.209		-5.0	
Hexachlorobenzene	0.255	0.238		-6.7	
Atrazine	0.228	0.228		0.0	
Pentachlorophenol	0.161	0.169		5.0	20.0
Phenanthrene	1.099	1.058		-3.7	
Anthracene	1.122	1.113		-0.8	
Carbazole	1.057	1.064		0.7	
Di-n-butylphthalate	1.300	1.314		1.1	
Fluoranthene	1.311	1.296		-1.1	20.0
Pyrene	1.300	1.255		-3.5	
Terphenyl-d14	1.093	1.041		-4.8	
Butylbenzylphthalate	0.589	0.607		3.1	
3,3-Dichlorobenzidine	0.497	0.497		0.0	
Benzo (a) anthracene	1.319	1.273		-3.5	
Chrysene	1.235	1.201		-2.8	
Bis(2-ethylhexyl)phthalate	0.867	0.866		-0.1	
Di-n-octyl phthalate	1.492	1.541		3.3	20.0
Benzo (b) fluoranthene	1.179	1.115		-5.4	
Benzo (k) fluoranthene	1.180	1.139		-3.5	
Benzo (a) pyrene	1.148	1.107		-3.6	20.0
Indeno (1,2,3-cd) pyrene	1.467	1.402		-4.4	
Dibenzo (a,h) anthracene	1.199	1.138		-5.1	
Benzo (g,h,i) perylene	1.178	1.141		-3.1	
1,2,4,5-Tetrachlorobenzene	0.578	0.548		-5.2	
1,4-Dioxane	0.472	0.461		-2.3	20.0
2,3,4,6-Tetrachlorophenol	0.377	0.381		1.1	

All other compounds must meet a minimum RRF of 0.010.