

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PORT06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2934</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/27/2025 09:39</u>
Lab File ID:	<u>BP025596.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.204	1.204		0.0	
Benzaldehyde	1.082	1.320		22.0	
Phenol-d6	1.544	1.563		1.2	
Phenol	1.620	1.636		1.0	20.0
bis(2-Chloroethyl)ether	1.263	1.242		-1.7	
2-Chlorophenol	1.359	1.356		-0.2	
2-Methylphenol	1.125	1.127		0.2	
2,2-oxybis(1-Chloropropane)	1.692	1.723		1.8	
Acetophenone	0.502	0.496		-1.2	
3+4-Methylphenols	1.560	1.513		-3.0	
n-Nitroso-di-n-propylamine	1.023	1.032	0.050	0.9	
Nitrobenzene-d5	0.395	0.395		0.0	
Hexachloroethane	0.563	0.563		0.0	
Nitrobenzene	0.353	0.352		-0.3	
Isophorone	0.700	0.708		1.1	
2-Nitrophenol	0.181	0.187		3.3	20.0
2,4-Dimethylphenol	0.312	0.310		-0.6	
bis(2-Chloroethoxy)methane	0.423	0.417		-1.4	
2,4-Dichlorophenol	0.310	0.310		0.0	20.0
Naphthalene	1.040	1.018		-2.1	
4-Chloroaniline	0.459	0.458		-0.2	
Hexachlorobutadiene	0.211	0.206		-2.4	20.0
Caprolactam	0.114	0.121		6.1	
4-Chloro-3-methylphenol	0.355	0.363		2.3	20.0
2-Methylnaphthalene	0.676	0.657		-2.8	
Hexachlorocyclopentadiene	0.349	0.280	0.050	-19.8	
2,4,6-Trichlorophenol	0.390	0.385		-1.3	20.0
2-Fluorobiphenyl	1.463	1.415		-3.3	
2,4,5-Trichlorophenol	0.427	0.424		-0.7	
1,1-Biphenyl	1.453	1.400		-3.6	
2-Chloronaphthalene	1.131	1.094		-3.3	
2-Nitroaniline	0.329	0.360		9.4	
Dimethylphthalate	1.465	1.499		2.3	
Acenaphthylene	1.871	1.827		-2.4	
2,6-Dinitrotoluene	0.309	0.331		7.1	
3-Nitroaniline	0.347	0.364		4.9	
Acenaphthene	1.098	1.046		-4.7	20.0
2,4-Dinitrophenol	0.165	0.204	0.050	23.6	
4-Nitrophenol	0.285	0.275	0.050	-3.5	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PORT06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2934</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/27/2025 09:39</u>
Lab File ID:	<u>BP025596.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.736	1.718		-1.0	
2,4-Dinitrotoluene	0.440	0.481		9.3	
Diethylphthalate	1.488	1.550		4.2	
4-Chlorophenyl-phenylether	0.681	0.690		1.3	
Fluorene	1.370	1.386		1.2	
4-Nitroaniline	0.356	0.391		9.8	
4,6-Dinitro-2-methylphenol	0.125	0.142		13.6	
n-Nitrosodiphenylamine	0.610	0.593		-2.8	20.0
2,4,6-Tribromophenol	0.269	0.296		10.0	
4-Bromophenyl-phenylether	0.220	0.217		-1.4	
Hexachlorobenzene	0.255	0.256		0.4	
Atrazine	0.228	0.234		2.6	
Pentachlorophenol	0.161	0.155		-3.7	20.0
Phenanthrene	1.099	1.066		-3.0	
Anthracene	1.122	1.113		-0.8	
Carbazole	1.057	1.047		-0.9	
Di-n-butylphthalate	1.300	1.367		5.2	
Fluoranthene	1.311	1.324		1.0	20.0
Pyrene	1.300	1.296		-0.3	
Terphenyl-d14	1.093	1.122		2.7	
Butylbenzylphthalate	0.589	0.630		7.0	
3,3-Dichlorobenzidine	0.497	0.494		-0.6	
Benzo (a) anthracene	1.319	1.299		-1.5	
Chrysene	1.235	1.231		-0.3	
Bis(2-ethylhexyl)phthalate	0.867	0.884		2.0	
Di-n-octyl phthalate	1.492	1.545		3.6	20.0
Benzo (b) fluoranthene	1.179	1.165		-1.2	
Benzo (k) fluoranthene	1.180	1.169		-0.9	
Benzo (a) pyrene	1.148	1.139		-0.8	20.0
Indeno (1,2,3-cd) pyrene	1.467	1.427		-2.7	
Dibenzo (a,h) anthracene	1.199	1.176		-1.9	
Benzo (g,h,i) perylene	1.178	1.148		-2.5	
1,2,4,5-Tetrachlorobenzene	0.578	0.565		-2.2	
1,4-Dioxane	0.472	0.463		-1.9	20.0
2,3,4,6-Tetrachlorophenol	0.377	0.394		4.5	

All other compounds must meet a minimum RRF of 0.010.