

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2934

Client ID : SSTDCCC040

Date Analyzed: 08/22/2025

Lab File ID: BP025544.D

Time Analyzed: 10:49

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	137158	7.778	591885	10.55	400784	14.39
UPPER LIMIT	274316	8.278	1183770	11.048	801568	14.889
LOWER LIMIT	68579	7.278	295943	10.048	200392	13.889
EPA SAMPLE NO.						
01 ENV-SW01	164036	7.78	640301	10.54	399889	14.40
02 ENV-SW02	160345	7.78	629322	10.54	401606	14.38
03 ENV-SW03	170537	7.78	719655	10.54	487572	14.39
04 ENV-SW04	193608	7.78	812188	10.55	530149	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.