

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2934

Client ID : SSTDCCC040

Date Analyzed: 08/25/2025

Lab File ID: BP025561.D

Time Analyzed: 12:13

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	176050	7.766	715069	10.54	441846	14.39
UPPER LIMIT	352100	8.266	1430140	11.037	883692	14.89
LOWER LIMIT	88025	7.266	357535	10.037	220923	13.89
EPA SAMPLE NO.						
01 PB169362BL	173347	7.78	665603	10.54	424932	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.