

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2934

Client ID: SSTDCCC040

Date Analyzed: 08/25/2025

Lab File ID: BP025561.D

Time Analyzed: 12:13

Instrument ID: BNA\_P

GC Column: ZB-GR ID: 0.25 (mm)

01

|                | IS4 (PHN)<br>AREA # | RT #   | IS5 (CRY)<br>AREA # | RT #  | IS6 (PRY)<br>AREA # | RT #   |
|----------------|---------------------|--------|---------------------|-------|---------------------|--------|
| 12 HOUR STD    | 900532              | 17.189 | 945335              | 21.63 | 1072630             | 25.001 |
| UPPER LIMIT    | 1801060             | 17.689 | 1890670             | 22.13 | 2145260             | 25.501 |
| LOWER LIMIT    | 450266              | 16.689 | 472668              | 21.13 | 536315              | 24.501 |
| EPA SAMPLE NO. |                     |        |                     |       |                     |        |
| PB169362BL     | 890876              | 17.19  | 1139400             | 21.63 | 1287990             | 25.00  |

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.