

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025543.D
Instrument ID: BNA_P

Contract: PORT06
SDG NO.: Q2934
DFTPP Injection Date: 08/22/2025
DFTPP Injection Time: 10:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	78.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.7 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025544.D	08/22/2025	10:49
ENV-SW01	Q2934-01	BP025553.D	08/22/2025	17:39
ENV-SW02	Q2934-02	BP025554.D	08/22/2025	18:20
ENV-SW03	Q2934-03	BP025555.D	08/22/2025	19:02
ENV-SW04	Q2934-04	BP025556.D	08/22/2025	19:43