

Surrogate Summary

SDG No.: Q2934

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2934-01	ENV-SW01	1,2-Dichloroethane-d4	50	57.6	115		70 (74)	130 (125)
		Dibromofluoromethane	50	49.9	100		70 (75)	130 (124)
		Toluene-d8	50	49.1	98		70 (86)	130 (113)
		4-Bromofluorobenzene	50	57.0	114		70 (77)	130 (121)
Q2934-02	ENV-SW02	1,2-Dichloroethane-d4	50	57.7	115		70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101		70 (75)	130 (124)
		Toluene-d8	50	48.4	97		70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.5	111		70 (77)	130 (121)
Q2934-03	ENV-SW03	1,2-Dichloroethane-d4	50	57.6	115		70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101		70 (75)	130 (124)
		Toluene-d8	50	48.6	97		70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.7	111		70 (77)	130 (121)
Q2934-04	ENV-SW04	1,2-Dichloroethane-d4	50	59.8	120		70 (74)	130 (125)
		Dibromofluoromethane	50	50.6	101		70 (75)	130 (124)
		Toluene-d8	50	48.4	97		70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.4	113		70 (77)	130 (121)
VX0822WBL01	VX0822WBL01	1,2-Dichloroethane-d4	50	57.9	116		70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101		70 (75)	130 (124)
		Toluene-d8	50	49.3	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.5	113		70 (77)	130 (121)
VX0822WBS01	VX0822WBS01	1,2-Dichloroethane-d4	50	52.6	105		70 (74)	130 (125)
		Dibromofluoromethane	50	48.3	97		70 (75)	130 (124)
		Toluene-d8	50	46.7	93		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.8	99		70 (77)	130 (121)
VX0822WBSD0	VX0822WBSD01	1,2-Dichloroethane-d4	50	49.6	99		70 (74)	130 (125)
		Dibromofluoromethane	50	47.9	96		70 (75)	130 (124)
		Toluene-d8	50	46.5	93		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.5	97		70 (77)	130 (121)

