

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

RunNo.: LB136914

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.490	0.50	98	90-110	08/21/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.493	0.5	99	90-110	08/21/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.505	0.5	101	90-110	08/21/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

RunNo.: LB136936

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	08/22/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	08/22/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	08/22/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	08/22/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	08/22/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

RunNo.: LB136944

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.500	0.50	100	90-110	08/25/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.495	0.50	99	90-110	08/25/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.497	0.50	99	90-110	08/25/2025