

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

Run No.: LB136914

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136914BS							
Orthophosphate as P	mg/L	0.5	0.51		102	1	90-110	08/21/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

Run No.: LB136925

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136925BS							
BOD5	mg/L	198	205		104	1	84.6-115.4	08/22/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

Run No.: LB136941

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136941BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	08/25/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

Run No.: LB136966

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136966BS							
TSS	mg/L	550	532		97	1	90-110	08/26/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

Run No.: LB136936

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169354BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	08/22/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2935

Project: Pre Treatment Plant 2025

Run No.: LB136944

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169384BS							
Phosphorus, Total	mg/L	0.50	0.50		99	1	90-110	08/25/2025