



CASE NARRATIVE

AECOM

Project Name: National Grid Equity - Brooklyn NY

Project # N/A Order ID # Q2936

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

5 Water samples were received on 08/21/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for MW-19B-082125 [1,2-Dichloroethane-d4 - 147%] Due to high concentration of compounds, this sample required dilution. Therefore, sample was reanalyzed with dilution and reported.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike for {VN0822WBS01} with File ID: VN087631.D met requirements for all compounds except for 1,2-Dichloropropane[81%] failing marginally low and associated sample required dilution therefore lab analyzed dilution under passing Blank Spike and reported.

The Blank analysis did not indicate the presence of lab contamination. The %RSD is greater than 20% in the Initial Calibration method (82N082125W.M) for Methylene Chloride passing on Linear regression.

The Continuous Calibration File ID VN087628.D met the requirements except for Acetone failing marginally low therefore no corrective action taken.

The Tuning criteria met requirements.



Samples MW-19B-082125, MW-19B-082125DL were diluted due to high concentrations.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

$Signature_{-}$			
Jignature_			