

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2936

Client ID : SSTDCCC040

Date Analyzed: 08/27/2025

Lab File ID: BP025596.D

Time Analyzed: 09:39

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	161611	7.778	681785	10.54	463868	14.40
UPPER LIMIT	323222	8.278	1363570	11.043	927736	14.896
LOWER LIMIT	80805.5	7.278	340893	10.043	231934	13.896
EPA SAMPLE NO.						
01 MW-19B-082125DL2	149853	7.78	592770	10.54	367926	14.40
02 MW-201-082025MS	177033	7.78	389081	10.57	432672	14.39
03 MW-201-082025MSD	165573	7.78	358938	10.57	417829	14.40
04 MW-19B-082125DL	151725	7.77	613134	10.54	392479	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.