

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025578.D
Instrument ID: BNA_P

Contract: AECO02
SDG NO.: Q2936
DFTPP Injection Date: 08/26/2025
DFTPP Injection Time: 08:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	78.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025579.D	08/26/2025	10:21
PB169362BS	PB169362BS	BP025581.D	08/26/2025	11:44
MW-19B-082125	Q2936-02	BP025591.D	08/26/2025	18:41