

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2936</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/25/2025 09:37</u>
Lab File ID:	<u>VN087647.D</u>	Init. Calib. Date(s):	<u>08/21/2025 08/21/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:09 14:44</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.710	0.739		4.09	20
Chloromethane	0.730	0.705	0.1	-3.42	20
Vinyl Chloride	0.846	0.778		-8.04	20
Bromomethane	0.437	0.430		-1.6	20
Chloroethane	0.595	0.545		-8.4	20
Trichlorofluoromethane	1.240	1.160		-6.45	20
1,1,2-Trichlorotrifluoroethane	0.701	0.637		-9.13	20
1,1-Dichloroethene	0.721	0.637		-11.65	20
Acetone	0.558	0.579		3.76	20
Carbon Disulfide	1.985	1.818		-8.41	20
Methyl tert-butyl Ether	2.704	2.628		-2.81	20
Methyl Acetate	1.168	1.177		0.77	20
Methylene Chloride	0.948	0.767		-19.09	20
trans-1,2-Dichloroethene	0.781	0.689		-11.78	20
1,1-Dichloroethane	1.546	1.366	0.1	-11.64	20
Cyclohexane	1.289	1.137		-11.79	20
2-Butanone	0.734	0.754		2.72	20
Carbon Tetrachloride	0.547	0.522		-4.57	20
cis-1,2-Dichloroethene	0.934	0.847		-9.31	20
Bromochloromethane	0.747	0.632		-15.4	20
Chloroform	1.578	1.444		-8.49	20
1,1,1-Trichloroethane	1.337	1.225		-8.38	20
Methylcyclohexane	0.610	0.578		-5.25	20
Benzene	1.699	1.596		-6.06	20
1,2-Dichloroethane	0.653	0.611		-6.43	20
Trichloroethene	0.388	0.345		-11.08	20
1,2-Dichloropropane	0.448	0.413		-7.81	20
Bromodichloromethane	0.653	0.623		-4.59	20
4-Methyl-2-Pentanone	0.713	0.738		3.51	20
Toluene	1.053	0.966		-8.26	20
t-1,3-Dichloropropene	0.676	0.666		-1.48	20
cis-1,3-Dichloropropene	0.702	0.693		-1.28	20
1,1,2-Trichloroethane	0.407	0.405		-0.49	20
2-Hexanone	0.451	0.505		11.97	20
Dibromochloromethane	0.472	0.454		-3.81	20
1,2-Dibromoethane	0.430	0.421		-2.09	20
Tetrachloroethene	0.347	0.307		-11.53	20
Chlorobenzene	1.244	1.143	0.3	-8.12	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2936</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/25/2025</u> <u>09:37</u>
Lab File ID:	<u>VN087647.D</u>	Init. Calib. Date(s):	<u>08/21/2025</u> <u>08/21/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:09</u> <u>14:44</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	2.154	2.040		-5.29	20
m/p-Xylenes	0.790	0.773		-2.15	20
o-Xylene	0.774	0.738		-4.65	20
Styrene	1.297	1.307		0.77	20
Bromoform	0.317	0.327	0.1	3.15	20
Isopropylbenzene	4.040	3.706		-8.27	20
1,1,2,2-Tetrachloroethane	1.391	1.318	0.3	-5.25	20
1,3-Dichlorobenzene	1.876	1.729		-7.84	20
1,4-Dichlorobenzene	1.952	1.749		-10.4	20
1,2-Dichlorobenzene	1.780	1.675		-5.9	20
1,2-Dibromo-3-Chloropropane	0.330	0.330		0	20
1,2,4-Trichlorobenzene	1.069	1.006		-5.89	20
1,2,3-Trichlorobenzene	1.045	0.985		-5.74	20
1,2-Dichloroethane-d4	1.024	0.883		-13.77	20
Dibromofluoromethane	0.361	0.318		-11.91	20
Toluene-d8	1.351	1.167		-13.62	20
4-Bromofluorobenzene	0.481	0.422		-12.27	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.