

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN082125**

|                          |                                                       |                   |                                |
|--------------------------|-------------------------------------------------------|-------------------|--------------------------------|
| Review By                | John Carlone                                          | Review On         | 8/22/2025 9:17:52 AM           |
| Supervise By             | Maresh Dadoda                                         | Supervise On      | 8/25/2025 8:11:38 AM           |
| SubDirectory             | VN082125                                              | HP Acquire Method | HP Processing Method 82N082125 |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                      |                   |                                |
| Tune/Reschk              | VP135214                                              |                   |                                |
| Initial Calibration Stds | VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 |                   |                                |
| CCC                      |                                                       |                   |                                |
| Internal Standard/PEM    |                                                       |                   |                                |
| ICV/I.BLK                | VP135221                                              |                   |                                |
| Surrogate Standard       |                                                       |                   |                                |
| MS/MSD Standard          |                                                       |                   |                                |
| LCS Standard             |                                                       |                   |                                |

| Sr# | Sampleld     | ClientID     | Data File Name | Date-Time         | Comment                              | Operator | Status |
|-----|--------------|--------------|----------------|-------------------|--------------------------------------|----------|--------|
| 1   | BFB          | BFB          | VN087614.D     | 21 Aug 2025 09:30 |                                      | JC\MD    | Ok     |
| 2   | VSTDCCC020   | VSTDCCC020   | VN087615.D     | 21 Aug 2025 10:04 | Need ICAL                            | JC\MD    | Not Ok |
| 3   | VN0821WBS01  | VN0821WBS01  | VN087616.D     | 21 Aug 2025 10:37 |                                      | JC\MD    | Not Ok |
| 4   | VN0821WBSD01 | VN0821WBSD01 | VN087617.D     | 21 Aug 2025 11:11 |                                      | JC\MD    | Not Ok |
| 5   | VN0821WBL01  | VN0821WBL01  | VN087618.D     | 21 Aug 2025 11:36 |                                      | JC\MD    | Not Ok |
| 6   | VSTDICC001   | VSTDICC001   | VN087619.D     | 21 Aug 2025 12:09 | LR-10,20                             | JC\MD    | Ok,M   |
| 7   | VSTDICC005   | VSTDICC005   | VN087620.D     | 21 Aug 2025 13:08 |                                      | JC\MD    | Ok,M   |
| 8   | VSTDICC020   | VSTDICC020   | VN087621.D     | 21 Aug 2025 13:41 | method not used for #N-amyle Acetate | JC\MD    | Ok,M   |
| 9   | VSTDICCC050  | VSTDICCC050  | VN087622.D     | 21 Aug 2025 14:02 |                                      | JC\MD    | Ok,M   |
| 10  | VSTDICC100   | VSTDICC100   | VN087623.D     | 21 Aug 2025 14:23 |                                      | JC\MD    | Ok,M   |
| 11  | VSTDICC150   | VSTDICC150   | VN087624.D     | 21 Aug 2025 14:44 |                                      | JC\MD    | Ok,M   |
| 12  | IBLK         | IBLK         | VN087625.D     | 21 Aug 2025 15:05 |                                      | JC\MD    | Ok     |
| 13  | VSTDICV050   | ICVVN082125  | VN087626.D     | 21 Aug 2025 15:47 | Icv Failed for Com.# 64 for DOD      | JC\MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN082225**

|              |                     |                   |                                |
|--------------|---------------------|-------------------|--------------------------------|
| Review By    | John Carlone        | Review On         | 8/25/2025 9:04:47 AM           |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 8/25/2025 3:20:26 PM           |
| SubDirectory | VN082225            | HP Acquire Method | HP Processing Method 82N082125 |

| STD. NAME                                                                                          | STD REF.#         |
|----------------------------------------------------------------------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135237          |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135240,VP135241 |

| Sr# | SampleID     | ClientID      | Data File Name | Date-Time         | Comment                                         | Operator | Status |
|-----|--------------|---------------|----------------|-------------------|-------------------------------------------------|----------|--------|
| 1   | BFB          | BFB           | VN087627.D     | 22 Aug 2025 08:00 |                                                 | JC\MD    | Ok     |
| 2   | VSTDCCC050   | VSTDCCC050    | VN087628.D     | 22 Aug 2025 09:16 | pH#Lot#V12668                                   | JC\MD    | Ok,M   |
| 3   | VN0822MBL01  | VN0822MBL01   | VN087629.D     | 22 Aug 2025 10:26 |                                                 | JC\MD    | Ok     |
| 4   | VN0822WBL01  | VN0822WBL01   | VN087630.D     | 22 Aug 2025 10:47 |                                                 | JC\MD    | Ok     |
| 5   | VN0822WBS01  | VN0822WBS01   | VN087631.D     | 22 Aug 2025 11:16 | BS failed low for comp.#45                      | JC\MD    | Ok,M   |
| 6   | VN0822WBSD01 | VN0822WBSD01  | VN087632.D     | 22 Aug 2025 11:55 | BSD failed Low for com.#45,57,60                | JC\MD    | Ok,M   |
| 7   | IBLK         | IBLK          | VN087633.D     | 22 Aug 2025 12:16 |                                                 | JC\MD    | Ok     |
| 8   | Q2920-02     | FB            | VN087634.D     | 22 Aug 2025 12:37 | vial A pH<2 FB;BS,BSD Failed Low                | JC\MD    | ReRun  |
| 9   | Q2920-01     | MW1           | VN087635.D     | 22 Aug 2025 12:58 | vial A pH<2 BS,BSD Failed Low                   | JC\MD    | ReRun  |
| 10  | Q2921-01     | MW1           | VN087636.D     | 22 Aug 2025 13:19 | vial A pH<2 BS,BSD Failed Low                   | JC\MD    | ReRun  |
| 11  | IBLK         | IBLK          | VN087637.D     | 22 Aug 2025 13:40 |                                                 | JC\MD    | Ok     |
| 12  | Q2936-05     | TB            | VN087638.D     | 22 Aug 2025 14:01 | vial A pH<2 TB;BS,BSD Failed Low                | JC\MD    | ReRun  |
| 13  | Q2936-04     | FB-02-082125  | VN087639.D     | 22 Aug 2025 14:22 | vial A pH<2 FB;BS,BSD Failed Low;Hit of com.#16 | JC\MD    | ReRun  |
| 14  | Q2936-01     | MW-19A-082125 | VN087640.D     | 22 Aug 2025 14:43 | vial A pH<2 BS,BSD Failed Low                   | JC\MD    | ReRun  |
| 15  | Q2924-06     | MW-2A-082025  | VN087641.D     | 22 Aug 2025 15:05 | Already Analyzed                                | JC\MD    | Not Ok |
| 16  | Q2936-03     | MW-19C-082125 | VN087642.D     | 22 Aug 2025 15:26 | vial A pH<2 BS,BSD Failed Low                   | JC\MD    | ReRun  |
|     |              |               |                |                   |                                                 |          |        |

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN082225**

|                                                                                                    |                     |                   |                      |  |           |  |  |
|----------------------------------------------------------------------------------------------------|---------------------|-------------------|----------------------|--|-----------|--|--|
| Review By                                                                                          | John Carlone        | Review On         | 8/25/2025 9:04:47 AM |  |           |  |  |
| Supervise By                                                                                       | Semsettin Yesilyurt | Supervise On      | 8/25/2025 3:20:26 PM |  |           |  |  |
| SubDirectory                                                                                       | VN082225            | HP Acquire Method | HP Processing Method |  | 82N082125 |  |  |
| STD. NAME                                                                                          |                     | STD REF.#         |                      |  |           |  |  |
| Tune/Reschk<br>Initial Calibration Stds                                                            |                     | VP135237          |                      |  |           |  |  |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard |                     | VP135240,VP135241 |                      |  |           |  |  |

|    |            |               |            |                   |                                                       |       |          |
|----|------------|---------------|------------|-------------------|-------------------------------------------------------|-------|----------|
| 17 | Q2936-02   | MW-19B-082125 | VN087643.D | 22 Aug 2025 15:47 | vial A pH<2 Surrogate fail;BS,BSD Failed Low;Need 20X | JC\MD | Dilution |
| 18 | Q2940-01   | COMP(1-6)     | VN087644.D | 22 Aug 2025 16:08 | vial A pH<2 E flag of previous sample                 | JC\MD | ReRun    |
| 19 | VSTDCCC050 | VSTDCCC050EC  | VN087645.D | 22 Aug 2025 16:41 |                                                       | JC\MD | Ok,M     |

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN082525**

|              |               |                   |                                   |
|--------------|---------------|-------------------|-----------------------------------|
| Review By    | John Carlone  | Review On         | 8/26/2025 9:08:17 AM              |
| Supervise By | Maresh Dadoda | Supervise On      | 8/26/2025 3:15:38 PM              |
| SubDirectory | VN082525      | HP Acquire Method | HP Processing Method 82N082125W.M |

| STD. NAME                                                                                          | STD REF.#         |
|----------------------------------------------------------------------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135285          |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135286,VP135287 |

| Sr# | SampleID     | ClientID         | Data File Name | Date-Time         | Comment               | Operator | Status   |
|-----|--------------|------------------|----------------|-------------------|-----------------------|----------|----------|
| 1   | BFB          | BFB              | VN087646.D     | 25 Aug 2025 08:15 |                       | JC\MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050       | VN087647.D     | 25 Aug 2025 09:37 | pH#Lot#V12668         | JC\MD    | Ok,M     |
| 3   | VN0825MBL01  | VN0825MBL01      | VN087648.D     | 25 Aug 2025 10:12 |                       | JC\MD    | Ok       |
| 4   | VN0825WBL01  | VN0825WBL01      | VN087649.D     | 25 Aug 2025 10:33 |                       | JC\MD    | Ok       |
| 5   | VN0825WBS01  | VN0825WBS01      | VN087650.D     | 25 Aug 2025 12:21 |                       | JC\MD    | Ok,M     |
| 6   | VN0825WBSD01 | VN0825WBSD01     | VN087651.D     | 25 Aug 2025 12:56 |                       | JC\MD    | Ok,M     |
| 7   | Q2936-02DL   | MW-19B-082125DL  | VN087652.D     | 25 Aug 2025 13:16 | vial B pH<2 Need 500X | JC\MD    | Dilution |
| 8   | Q2936-05     | TB               | VN087653.D     | 25 Aug 2025 13:37 | vial B pH<2 TB        | JC\MD    | Ok       |
| 9   | Q2936-04     | FB-02-082125     | VN087654.D     | 25 Aug 2025 13:58 | vial B pH<2 FB        | JC\MD    | Ok       |
| 10  | Q2920-02     | FB               | VN087655.D     | 25 Aug 2025 14:19 | vial B pH<2 FB        | JC\MD    | Ok       |
| 11  | Q2936-03     | MW-19C-082125    | VN087656.D     | 25 Aug 2025 14:40 | vial B pH<2           | JC\MD    | Ok       |
| 12  | Q2936-01     | MW-19A-082125    | VN087657.D     | 25 Aug 2025 15:01 | vial B pH<2           | JC\MD    | Ok       |
| 13  | Q2920-01     | MW1              | VN087658.D     | 25 Aug 2025 15:22 | vial B pH<2           | JC\MD    | Ok       |
| 14  | Q2921-01     | MW1              | VN087659.D     | 25 Aug 2025 15:43 | vial B pH<2           | JC\MD    | Ok       |
| 15  | PB169356TB   | PB169356TB       | VN087660.D     | 25 Aug 2025 16:04 |                       | JC\MD    | Ok       |
| 16  | Q2936-02DL2  | MW-19B-082125DL2 | VN087661.D     | 25 Aug 2025 16:25 | vial C pH<2           | JC\MD    | Ok       |
| 17  | Q2941-02     | SOIL-COMP(1-2)   | VN087662.D     | 25 Aug 2025 16:46 | vial A pH#5.0         | JC\MD    | Ok       |
| 18  | VN0825WBSD02 | VN0825WBSD02     | VN087663.D     | 25 Aug 2025 17:08 | Not Required          | JC\MD    | Not Ok   |

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN082525**

|              |               |                   |                                   |
|--------------|---------------|-------------------|-----------------------------------|
| Review By    | John Carlone  | Review On         | 8/26/2025 9:08:17 AM              |
| Supervise By | Maresh Dadoda | Supervise On      | 8/26/2025 3:15:38 PM              |
| SubDirectory | VN082525      | HP Acquire Method | HP Processing Method 82N082125W.M |

| STD. NAME                                                                                          | STD REF.#         |
|----------------------------------------------------------------------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135285          |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135286,VP135287 |

|    |            |              |            |                   |                   |       |        |
|----|------------|--------------|------------|-------------------|-------------------|-------|--------|
| 19 | Q2909-02   | 4065         | VN087664.D | 25 Aug 2025 17:29 | Need Straight Run | JC\MD | Not Ok |
| 20 | Q2940-01   | COMP(1-6)    | VN087665.D | 25 Aug 2025 17:49 | vial B pH<2       | JC\MD | Ok     |
| 21 | VSTDCCC050 | VSTDCCC050EC | VN087666.D | 25 Aug 2025 18:10 |                   | JC\MD | Ok,M   |

M : Manual Integration