

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q2936
 Lab File ID: VN087628.D Date Analyzed: 08/22/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:16
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	263551	8.21	512519	9.08	469406	11.85
UPPER LIMIT	527102	8.706	1025040	9.582	938812	12.347
LOWER LIMIT	131776	7.706	256260	8.582	234703	11.347
EPA SAMPLE NO.						
MW-19B-082125	165286	8.21	419366	9.08	365004	11.85
VN0822WBL01	212794	8.21	479771	9.08	448296	11.84
VN0822WBS01	241192	8.21	479192	9.08	431937	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.