

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2937

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB136926

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136926BS							
TDS	mg/L	100	95.0		95	1	90-110	08/22/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2937

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB136966

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136966BS							
TSS	mg/L	550	532		97	1	90-110	08/26/2025