



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** Tetra Tech NUS, Inc.

**SDG No.:** Q2937

**Contract:** TETR06

**Lab Code:** ACE

**Initial Calibration Source:** \_\_\_\_\_

**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|--------------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CRI01     | Iron    | 96.8           | 100                | 97            | 65 - 135                  | P | 08/26/2025       | 11:29            | LB136973      |