

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc.

SDG No.: Q2937

Contract: TETR06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	99200	101000	98	85600	116500	08/26/2025	11:33	LB136973
ICSAB01	Iron	100000	99300	101	84400	114500	08/26/2025	11:37	LB136973
ICSA	Iron	95300	101000	94	85600	116500	08/26/2025	11:41	LB136973
ICSAB	Iron	96500	99300	97	84400	114500	08/26/2025	11:45	LB136973